

CALIBRATION CERTIFICATE

Total Weighing Solutions Ltd
Unit 8 Signal Park
Daventry
Northamptonshire
NN11 4RU
Tel: 01327 368020
www.totalweighing.co.uk



Customer Details

Name: WHILTON MILL LTD
Address: WHILTON LOCKS

DAVENTRY
NORTHAMTONSHIRE
NN1 2NH
Contact: DAN
Telephone: 01327 843822


Linearity / Accuracy Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0.0	0.0	0.0	0.0	0.0
60.0	60.0	0.0	60.0	0.0
120.0	120.1	0.1	120.0	0.0
200.0	200.2	0.2	200.0	0.0
300.0	300.4	0.4	300.0	0.0

Certificate No. 20676

Date of Calibration: 26/11/252025
Next Due: 26/11/262026
Sales Order:

Diagram



Scale Details

Location:
Manufacturer: DINI ARGEO
Model No.: DFWLB-2
Serial No.: 722941484
Capacity: 600 Kg
Resolution: 0.1
Tolerance: +/- 3 divs
Frequency: Yearly

Off Centre Load Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0

Test Weight/Rig Details

Identification No. and Certificate No.
Test Weights and Equipment are traceable to National Standards

WEIGHT SET:
T1 - T35, D11724 & B12267

CERTIFICATE NUMBER:
13083, 11470, 11471 & 121466

Repeatability Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
300.0	300.0	0.0	300.0	0.0
300.0	300.0	0.0	300.0	0.0
300.0	300.0	0.0	300.0	0.0

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments

Comments

Tolerances

Unless otherwise agreed, tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows:-
0-500 divisions +/- 1div
>501-2000 divisions +/- 2div
>2001 divisions +/- 3div

Technicians Name: Tony Goodman

Technician Signature:

