


TopWeigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK



Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

| | | | | | | |
|--|---------|------------------|---------|---|--|-------------------|
| Customer Cumbria Kart Racing Club | | | | Customer Order No. | | |
| | | | | Date of Last Test | | 3 Feb 2024 |
| | | | | Date of Test | | 1 Feb 2025 |
| | | | | Next Test Due | | 1 Feb 2026 |
| Site Address :- Rowrah Cumbria | | | | Next Test Due within 30 days of above. | | |
| | | | | Machine | | Platform |
| | | | | Type | | |
| | | | | Manufacturer | | TopWeigh Ltd |
| | | | | Capacity :- Max | | 300 kg |
| | | | | Min | | |
| | | | | e. | | 0.1 kg |
| Calibration Certificate No. | | TW | | 1025127 | | Maximum Tolerance |
| Initial Test | | Calibration Test | | Serial No. | | JWS-06-300/1 |
| Weights Applied | Reading | Weights Applied | Reading | Class | | (111) |
| | | | | Observations & Comments :- | | |
| | | | | Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT | | |
| | | | | Weights Traceable to National Standards TW 11985 | | |
| | | | | Weights Used TWR Set | | |

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :- *A J Jeffrey* Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.



TopWeigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK




Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate


| | | | | | |
|--|---------|------------------|---------|---|---------------|
| Customer Cumbria Kart Racing Club | | | | Customer Order No. | |
| | | | | Date of Last Test | 3 Feb 2024 |
| | | | | Date of Test | 1 Feb 2025 |
| | | | | Next Test Due | 1 Feb 2026 |
| Site Address :- Rowrah Cumbria | | | | Next Test Due within 30 days of above. | |
| | | | | Machine | Dail Platform |
| | | | | Type | Avery |
| | | | | Manufacturer | 4208 GLE |
| | | | | Capacity :- Max | 250 kg |
| | | | | Min | |
| | | | | e. | 500g |
| Calibration Certificate No. | | TW | | 1025128 | |
| Initial Test | | Calibration Test | | Maximum Tolerance | |
| Serial No. | | B603951 | | | |
| Weights Applied | Reading | Weights Applied | Reading | | |
| | | | | | |
| | | | | Observations & Comments :- | |
| | | | | Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 | |
| | | | | Found to be CORRECT | |
| | | | | | |
| | | | | | |
| | | | | Weights Traceable to National Standards | |
| | | | | TW 11985 | |
| | | | | Weights Used | |
| | | | | TWR Set | |

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :-  Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.



TopWeigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK



Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
Mobile : +44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

| | | | | |
|--|---------|--|---------|---|
| Customer Cumbria Kart Racing Club | | Customer Order No. | | |
| | | Date of Last Test | | 3 Feb 2024 |
| | | Date of Test | | 1 Feb 2025 |
| | | Next Test Due | | 1 Feb 2026 |
| Site Address :- Rowrah Cumbria | | Next Test Due within 30 days of above. | | |
| | | Machine | | |
| | | Type | | Test Weights |
| | | Manufacturer | | |
| | | Capacity :- Max | | |
| | | Min | | |
| | | e. | | |
| Calibration Certificate No. | | TW | 1025129 | Maximum Tolerance |
| Initial Test | | Calibration Test | | Serial No. |
| | | | | RCC2-RCC6 |
| Weights Applied | Reading | Weights Applied | Reading | |
| RCC 2 | 20 Kg | | | Observations & Comments :- Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT |
| RCC 3 | 20 Kg | | | |
| RCC 4 | 20 Kg | | | |
| RCC 5 | 20 Kg | | | |
| RCC 6 | 20 Kg | | | |
| | | | | Weights Traceable to National Standards TW 11985 |
| | | | | Weights Used TWR Set |

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :- Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.