



TopWeigh Ltd
 Scale House, (Jeffrey Estates), Rockcliffe,
 Carlisle, Cumbria. CA6 4BH. UK




Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
 Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

Customer Cumbria Kart Racing Club		Customer Order No.		
		Date of Last Test		1 Feb 2025
		Date of Test		21 Feb 2026
		Next Test Due		21 Feb 2027
Site Address :- Rowrah Cumbria		Next Test Due within 30 days of above.		
		Machine		Platform
		Type		
		Manufacturer		TopWeigh Ltd
		Capacity :- Max		300 kg
		Min		
		e.		0.1 kg
Calibration Certificate No.		TW	1026111	Maximum Tolerance
Initial Test		Calibration Test		Serial No.
		JWS-06-300/1		
Weights Applied	Reading	Weights Applied	Reading	Class
				(111)
Observations & Comments :- Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT				
Weights Traceable to National Standards TW 11985				
Weights Used TWR Set				

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :-  Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.



TopWeigh Ltd

**Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK**



Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
 Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate


Customer Cumbria Kart Racing Club		Customer Order No.		
		Date of Last Test		1 Feb 2025
		Date of Test		21 Feb 2026
		Next Test Due		21 Feb 2027
Site Address :- Rowrah Cumbria		Next Test Due within 30 days of above.		
		Machine		Dail Platform
		Type		Avery
		Manufacturer		4208 GLE
		Capacity :- Max		250 kg
		Min e.		500g
Calibration Certificate No.		TW	1026112	
Initial Test		Calibration Test		Serial No.
		B603951		
Weights Applied	Reading	Weights Applied	Reading	
				Observations & Comments :- Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT
				Weights Traceable to National Standards TW 11985
				Weights Used TWR Set

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :- Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.



TopWeigh Ltd
 Scale House, (Jeffrey Estates), Rockcliffe,
 Carlisle, Cumbria. CA6 4BH. UK




Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
 Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

Customer Cumbria Kart Racing Club		Customer Order No.					
		Date of Last Test					
		Date of Test		26 Mar 2026			
		Next Test Due		26 Mar 2027			
Site Address :- Rowrah Cumbria		Next Test Due within 30 days of above.					
		Machine		Platform			
		Type					
		Manufacturer		TopWeigh Ltd			
		Capacity :- Max		300 kg			
		Min e.		0.1 kg			
Calibration Certificate No.		TW	1026117				
Initial Test		Calibration Test		Serial No.			
		0742720666					
Weights Applied	Reading	Weights Applied	Reading	Class			
				(111)			
				Observations & Comments :-			
				Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT			
				Weights Traceable to National Standards			
				TW 11985			
Weights Used				TWR Set			

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :-  Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.