

CALIBRATION CERTIFICATE

Total Weighing Solutions Ltd
Unit 5 Alvis Way, Royal Oak Industrial Estate
Daventry, NN11 8PG, UK
www.totalweighing.co.uk
Tel: 01327 368020

Customer Details

Name: WHILTON MILL LTD
Address: WHILTON LOCKS
DAVENTRY
NORTHANTS
NN1 2NH
Contact: DAN
Telephone:

Certificate No. 14140

Date of Calibration: 01/12/2020

Next Due: 01/12/2021

Location:
Department:
Manufacturer: DINI ARGEO
Model: DFWLB-2
Serial No.: 0100482844
Capacity: 600 Kg
Resolution: 0.1
Tolerance: +/- 3 divs
Calibration Frequency: YEARLY

Test Weight/Rig Details

Identification No. and Certificate No.

WEIGHT SET:
TWS1 - TWS17, D11724 & B12267

CERTIFICATE NUMBER:
8429, 8430, 8431 & 8432


Test weights and equipment are traceable to ISO17025

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments

Linearity / Accuracy Test

Test Point Kg	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0	0.0	0.0	0.0	0.0
40	40.0	0.0	120	80.0
120	119.9	-0.1	119.9	-0.1
160	160.0	0.0	160.0	0.0
200	199.9	-0.1	199.9	-0.1

Diagram



Off Centre Load Test

Test Point Kg	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
40	40.0	0.0	40.0	0.0
40	40.0	0.0	40.0	0.0
40	39.9	-0.1	39.9	-0.1
40	40.0	0.0	40.0	0.0

Repeatability Test

Test Point Kg	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
120	120.0	0.0	120.0	0.0
120	120.0	0.0	120.0	0.0
120	120.0	0.0	120.0	0.0

Comments

Tolerances

Unless otherwise agreed, tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows:-
0-500 divisions +/- 1 div.
>501-2000 divisions +/- 2 div
>2001 divisions +/- 3 div

Technicians Name Ken Goodman

Technicians Signature